

DEVI

The surgical gloves are intended to act as a protective barrier between the patient and health care worker during invasive surgeries and also to facilitate general hand hygiene.

DEVICE CATALOG NO.:

Variants	Non-Sterile	Brand Name
Latex Surgical Gloves	SRGLPN	SAFEX GLOVES
Elbow length Gynecology gloves	ELGLPN	
Double pair gloves (High risk)	DPGLPN	
Under gloves	SUGLPN	
Microsurgery Gloves	MICLPN	
Orthopedic gloves	ORTLPN	
Ultra gloves Beaded	ULTLPNBD	
Ultra Gloves Beadles	ULTLPNBL	

DEVICE SIZES:

Variants	Sizes
Latex Surgical Gloves	5.5,6.0,6.5,7.0,7.5,8.0,8.5,9.0
Elbow length Gynecology gloves	6.5, 7.5, 8.5
Double pair gloves (High risk)	Inner:6.0,6.5,7.0,7.5,8.0,8.5, 9.0 Outer: 5.5,6.0,6.5,7.0,7.5,8.0,8.5
Under gloves	5.5,6.0,6.5,7.0,7.5,8.0,8.5,9.0
Microsurgery Gloves	5.5,6.0,6.5,7.0,7.5,8.0,8.5,9.0
Orthopedic gloves	5.5,6.0,6.5,7.0,7.5,8.0,8.5,9.0
Ultra gloves Beaded	5.5,6.0,6.5,7.0,7.5,8.0,8.5,9.0
Ultra Gloves Beadless	5.5,6.0,6.5,7.0,7.5,8.0,8.5,9.0

DEVICE DESCRIPTION:

Latex Surgical Gloves – Non-Sterile (Powdered Gloves)

Latex Surgical gloves are Non-Sterile anatomically shaped medical gloves with the thumb positioned towards the palmar surface of the index finger rather than lying flat. These gloves are Pre- powdered. The surgical gloves are intended for use in invasive surgeries.

Elbow length Gynaecology Procedure Latex Gloves- Non-Sterile (Powdered Gloves)

Elbow Length gynecology Gloves anatomically shaped medical gloves with the thumb positioned towards the palmar surface of the index finger rather than lying flat. These gloves are pre-powdered. The Elbow length gynecology gloves are intended for use various gynecological procedures.

Double Pair Gloves (High risk gloves) – Non-Sterile (Powdered Gloves)

The double paired sterile gloves are anatomically shaped medical gloves with the thumb positioned towards the palmar surface of the index finger rather than lying flat. The double pair specialty gloves are supplied in pairs to protect healthcare from contiguous disease. The inner glove is supplied green compared to the outer glove which helps users to notice any perforation on the outer gloves.

Under Gloves – Non-Sterile (Powdered Gloves)

Under gloves provide superior feel and natural sensitivity to the fingertips.

Microsurgery Gloves – Non-Sterile (Powdered Gloves)

Microsurgery gloves provide superior feel and natural sensitivity to the fingertips during microsurgery. The microsurgery gloves are intended for micro surgery like ophthalmic surgeries.

Orthopaedic Gloves – Non-Sterile (Powdered Gloves)

Orthopedic gloves anatomically shaped medical gloves with the thumb positioned towards the palmar surface of the index finger rather than lying flat. Its brown color gloves. These gloves are intended for orthopedic surgery.

Ultra Gloves Beaded – Non-Sterile (Powdered Gloves)

Ultra Gloves anatomically shaped medical gloves with the thumb positioned towards the palmar surface of the index finger rather than lying flat and Intended for surgery.

Ultra Gloves Beadless – Non- Sterile (Powdered Gloves)

Ultra Gloves anatomically shaped medical gloves with the thumb positioned towards the palmar surface of the index finger rather than lying flat and Intended for surgery.

INTENDED USER:

This is used by surgeons, operation theater staff and healthcare professional

TARGET PATIENT POPULATION:

Used for all patient populations except in patients with known allergy to natural latex rubber.

Rev. No.	A	Rev. Date	14-08-2022	Doc No:	BEO/IFU/SR G- PD/NS	Page 2 of 6	Device Name	Latex Surgical Examination Gloves, Powdered
----------	---	-----------	------------	---------	---------------------	-------------	-------------	---

DEVICE SHELF LIFE:

5 years from the date of manufacture printed on the Product Label

STORAGE INSTRUCTION:

It is recommended to store the gloves in a dry place, in the temperature of 5°C to 40°C and to protect them against direct sunlight.

METHOD OF STERILIZATION:

Supplied non sterile, suggested to be sterilized using Ethylene oxide before using.

INDICATIONS:

Intended to be worn on the hands during medical surgical procedures to provide a barrier against potentially infectious fluids and other contaminants.

WARNINGS AND PRECAUTIONS:

1. Do not use after the expiry date printed on the label.
2. Do not use it if the Package is opened or damaged.
3. The supplied gloves are non-sterile and suggested to be sterilized using EO before use.
4. Do not Reuse - The reuse of the gloves may cause severe complications such as cross contamination, dermatitis, loss of physical properties, infection, allergic reaction and poor barrier protection.
5. Latex gloves are made of Natural rubber latex, which may cause allergic reactions including Latex Allergy – Anaphylactic if the user is allergic to latex.
6. Single Use Packaging label and Sales Packaging Label.
7. Device should not be stored in direct sunlight

CONTRA INDICATIONS:

1. Gloves contain Powder content, persons who are sensitive to powder may develop allergy towards the powder content
2. Latex gloves are made of Natural rubber latex, which may cause allergic reactions including Latex Allergy – Anaphylactic if the user is allergic to latex.
3. Gloves contain Natural Latex; persons who are sensitive to Latex should consult a physician before using

DIRECTION FOR USE:

- Select the appropriate size Gloves for the hand.
- Take out the Gloves from the pouch.
- Wear the gloves as left to the Left hand & Right to the Right Hand as denoted on the wallet.
- Must check the date of manufacturing and expiry date before use.
- Adjust the gloves as needed.

Rev. No.	A	Rev. Date	24-6-2022	Doc No:	MRK/IFU/SR G- PD/NS	Page 3 of 6	Device Name	Latex Surgical Examination Gloves, Powdered
----------	---	-----------	-----------	---------	------------------------	----------------	-------------	--

A) How to Donning Gloves:

- Rubbing the hand with alcohol or wash with water
- Dry hands before donning the gloves.
- Take out gloves from single unit packaging.
- Gloves need to be donned carefully; do not apply excessive force to pull the gloves when donning.
- Keep the fingernails short and change gloves frequently to avoid tearing issue.
- For Double Pair Gloves, first wear the green color inner glove & then wear the outer glove over it.

B) How to Remove Gloves:

- The gloves should be discarded with an appropriate waste disposal container after use according to local disposal regulations

DURATION OF USE:

Duration of Use is short term use (< 24 hours)

MEDICAL BENEFITS:

- NRL gloves are a competent barrier to protect against infections for both healthcare professionals and the patients.
- NRL gloves provide lower rates of perforation and lower viral leakage rates.
- NRL gloves are easy to put on, comfortable to wear and provide adequate, durable protection.
- NRL gloves have good barrier integrity.
- NRL gloves have less after-use defects.
- NRL Gloves has significant greater satisfaction with regard to factors such as quality, safety and durability.
- NRL gloves have high tear propagation strength
- NRL gloves have low perforation rate
- NRL gloves have high tensile strength
- NRL gloves have good fit and comfort

SIDE EFFECTS:

Latex Allergy, Surgical site infection, Inflammation, Allergy, Pain, Rashes, Itching, Peeling of skin, Hypersensitivity type-I reaction, Skin redness, Ulcerated skin

RESIDUAL RISK:

- Anaphylaxis
- Blood-Borne Infection, Post-operative wound infection
- Respiratory tract & Pulmonary diseases
- Inflammation
- Toxic to environment










Rev. No.	A	Rev. Date	14-08-2022	Doc No:	BEO/IFU/SR G- PD/NS	Page 4 of 6	Device Name	Latex Surgical Examination Gloves Powdered
----------	---	-----------	------------	---------	---------------------	-------------	-------------	--









DEVICE SINGLE UNIT PACKAGING (PRIMARY PACK) CONTAINS:

All Variants	1 Pair/02 Pcs
Double pair gloves (High risk)	Inner Gloves – 1 Pair /02 Pcs Outer Gloves – 1 Pair/02 Pcs

DEVICE SALES PACKAGING (SECONDARY PACK) CONTAINS:

Variants	Each Pack Contains
Latex Surgical Gloves	50 Pairs
Elbow length Gynecology gloves	25 Pairs
Double pair gloves (High risk)	20 Pairs
Under gloves	40 Pairs
Microsurgery Gloves	40 Pairs
Orthopedic gloves	40 Pairs
Ultra gloves Beaded	40 Pairs
Ultra Gloves Beadles	40 Pairs

SYMBOLS USED IN DEVICE SINGLE UNIT PACKAGING AND SALES PACKAGING					
	Manufacturer		Date of Manufacture		Keep away from Sunlight
	Authorized Representative in the European Community/ Authorized representative European Union		Use by Date/ Expiry Date		Keep Dry
	Contains presence of Natural Rubber Latex		Batch Code		Use only once

	No: of gloves in the package		Do not use if package is damaged and consult instructions for use		Caution
			Medical Device		Unique Device Identifier
	Non sterile symbol		CE Mark Symbol		Storage Temperature

- In case of any adverse event noted, contact manufacturer (Blinkex Overseas) or authorized representative and competent authority of the member state where user/ patient is established without any delay or within 24 hrs



Blinkex Overseas

S.NO : 4520

Venkatesh Manor , Degloor , Nanded , Maharashtra , India - 431717